

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 08/22/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:42
INTERNAL STANDARD	VERIFIED	04:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:42
BLANK TEST	0.00	04:43
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:43
BLANK TEST	0.00	04:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:44
BLANK TEST	0.00	04:45



565 *Ralph J Bartlett*

08-22-11