

SOLUTION CHANGE

-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206  
 -----

Date: 08/29/2011

Operator Name: JOSHUA M LOPEZ  
 Permit #: 34728  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:05
INTERNAL STANDARD	VERIFIED	18:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:05
BLANK TEST	0.00	18:06
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	18:06
BLANK TEST	0.00	18:07
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	18:07
BLANK TEST	0.00	18:08

*JM Lopez* #299  
 8/29/11

