

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 09/13/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:38
INTERNAL STANDARD	VERIFIED	04:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:38
BLANK TEST	0.00	04:39
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:39
BLANK TEST	0.00	04:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:40
BLANK TEST	0.00	04:41



SIR *Ralph J Bartlett* WSO #162 09-13-11