

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 09/27/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:47
INTERNAL STANDARD	VERIFIED	02:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:48
BLANK TEST	0.00	02:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:49
BLANK TEST	0.00	02:50
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:50
BLANK TEST	0.00	02:51



SGS *Ralph Bartlett*

WCSO SHERIFF #102 09-27-11