

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/02/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:55
INTERNAL STANDARD	VERIFIED	02:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:56
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:58



515 Ralph Bartlett

WESO #162

10-2-11