

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/10/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:12
INTERNAL STANDARD	VERIFIED	23:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	23:13
BLANK TEST	0.00	23:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:14
BLANK TEST	0.00	23:15
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:15
BLANK TEST	0.00	23:16



505 Ralph Bartlett WCSO #162 10/10/11