

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206




Date: 10/20/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:24
INTERNAL STANDARD	VERIFIED	03:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:24
BLANK TEST	0.00	03:25
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:25
BLANK TEST	0.00	03:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:26
BLANK TEST	0.00	03:27

@@▲?Ç◆◎*@Ç▼0



SG *Ralph Bartlett* WSC #162 10/20/11