

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/25/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:52
INTERNAL STANDARD	VERIFIED	22:52
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:53
BLANK TEST	0.00	22:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:54
BLANK TEST	0.00	22:55



SGS Ralph Bartlett

INSTR #112

10-25-11