

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 12/02/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:56
INTERNAL STANDARD	VERIFIED	03:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:57
BLANK TEST	0.00	03:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:58
BLANK TEST	0.00	03:59



SGT Rust J Bartlett

WCSO Sheriff #162 12-02-11