

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/10/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:58
INTERNAL STANDARD	VERIFIED	02:58
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:58
BLANK TEST	0.00	02:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:01
BLANK TEST	0.00	03:01



Sgt. Russel Brown WISO #162 12-10-11