

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/15/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:17
INTERNAL STANDARD	VERIFIED	02:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:19
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:20
BLANK TEST	0.00	02:21



SGS Rush 1 Bartlett

WCSO #102

12-15-11