

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 12/25/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:41
INTERNAL STANDARD	VERIFIED	20:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:41
BLANK TEST	0.00	20:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:42
BLANK TEST	0.00	20:43
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	20:43
BLANK TEST	0.00	20:44



SGS Rush / Burt 12-25-11 w/s 10 #162