

SOLUTION CHANGE




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206  
 -----

Date: 12/30/2011

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:54
INTERNAL STANDARD	VERIFIED	03:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:54
BLANK TEST	0.00	03:55
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	03:55
BLANK TEST	0.00	03:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57



505 R. J. Bartlett