

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 01/07/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:02
INTERNAL STANDARD	VERIFIED	03:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:03
BLANK TEST	0.00	03:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:04
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:05
BLANK TEST	0.00	03:06



S65 *Ralph Bartlett* WORO #162 01-07-12