

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206  
 -----

Date: 01/13/2012

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:30
INTERNAL STANDARD	VERIFIED	02:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:31
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:32
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:33
BLANK TEST	0.00	02:33



SCJ *Ralph Bartlett* WESO #162 01-13-12