

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/01/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:36
INTERNAL STANDARD	VERIFIED	02:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:37
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:39

