

SUPERVISOR TEST

 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 02/09/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:20
INTERNAL STANDARD	VERIFIED	01:20
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:21
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:23

