

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/29/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:05
INTERNAL STANDARD	VERIFIED	03:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:06
BLANK TEST	0.00	03:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:07
BLANK TEST	0.00	03:07
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:08
BLANK TEST	0.00	03:09



Handwritten signature and notes at the bottom of the page.