

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/08/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:29
INTERNAL STANDARD	VERIFIED	02:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:30
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:31
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:32
BLANK TEST	0.00	02:33

