

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206  
 -----

Date: 03/22/2012

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:48
INTERNAL STANDARD	VERIFIED	04:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:49
BLANK TEST	0.00	04:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:50
BLANK TEST	0.00	04:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:51
BLANK TEST	0.00	04:51



SGT *Ralph Bartlett* WESO SHERIFF #162 03-22-12