

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/28/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:51
INTERNAL STANDARD	VERIFIED	02:52
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:52
BLANK TEST	0.00	02:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55



SGS RALPH J BARTLETT WESO #162 03-28-12