

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 04/10/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:00
INTERNAL STANDARD	VERIFIED	03:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:01
BLANK TEST	0.00	03:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:02
BLANK TEST	0.00	03:03



SGT Ralph J Bartlett WESO

SHERIFF #166 04-10-12