

SUPERVISOR TEST




WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206

Date: 04/25/2012

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:09
INTERNAL STANDARD	VERIFIED	03:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:10
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:11
BLANK TEST	0.00	03:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:12
BLANK TEST	0.00	03:12



505 Ralph J Bartlett

WCSO #162

04-25-12