

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 04/30/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:27
INTERNAL STANDARD	VERIFIED	03:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:27
BLANK TEST	0.00	03:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:28
BLANK TEST	0.00	03:29
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:29
BLANK TEST	0.00	03:30



565 Ralph Bartlett WSCSO Sheriff #162 04-30-12