

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/08/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:05
INTERNAL STANDARD	VERIFIED	01:06
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:08
BLANK TEST	0.00	01:09

