

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/14/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:00
INTERNAL STANDARD	VERIFIED	04:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:00
BLANK TEST	0.00	04:01
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:01
BLANK TEST	0.00	04:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:02
BLANK TEST	0.00	04:03



cos R J B

2-1-12