

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/23/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:21
INTERNAL STANDARD	VERIFIED	02:21
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:23
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:24
BLANK TEST	0.00	02:25



SGT Ralph Bartlett WESO #162

05-23-12