

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/31/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:41
INTERNAL STANDARD	VERIFIED	02:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:42
BLANK TEST	0.00	02:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:43
BLANK TEST	0.00	02:44
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:44
BLANK TEST	0.00	02:45



SGT *Ralph Bartlett* WESO #162

05/31/12