

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 06/06/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:12
INTERNAL STANDARD	VERIFIED	06:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:12
BLANK TEST	0.00	06:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:13
BLANK TEST	0.00	06:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:14
BLANK TEST	0.00	06:15



*SGT Ralph J Bartlett* 6/6/12 WESO #162