

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 06/14/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:34
INTERNAL STANDARD	VERIFIED	02:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:34
BLANK TEST	0.00	02:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:35
BLANK TEST	0.00	02:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:37
BLANK TEST	0.00	02:37



505 Ralph J Bartlett WCSO #162 06-14-12