

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 06/20/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:23
INTERNAL STANDARD	VERIFIED	01:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:27



565 Ralph J Bartlett wco #162 06-20-12