

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/28/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:00
INTERNAL STANDARD	VERIFIED	01:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:02
BLANK TEST	0.00	01:03



567 Ralph J Bartlett 06-28-12 WESO #162