

SUPERVISOR TEST

 WARREN CO.
 SHERIFF'S OFFICE
 DATAWASTER dmt: 106206

 Date: 07/25/2012

Operator Name: JOSHUA M LOPEZ
 Permit #: 34728
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:43
INTERNAL STANDARD	VERIFIED	20:43
REFERENCE STANDARD	33.99C	
REFERENCE STANDARD	0.09	20:43
BLANK TEST	0.00	20:44
REFERENCE STANDARD	34.02C	
REFERENCE STANDARD	0.09	20:44
BLANK TEST	0.00	20:45
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.09	20:45
BLANK TEST	0.00	20:46

J.M. Lopez
 #299

