

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/09/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:25
INTERNAL STANDARD	VERIFIED	03:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:26
BLANK TEST	0.00	03:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:27
BLANK TEST	0.00	03:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:28
BLANK TEST	0.00	03:29



SGS *Ralph J Bartlett* 08-09-12 WCSO #162