

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/15/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:16
INTERNAL STANDARD	VERIFIED	02:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:16
BLANK TEST	0.00	02:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:17
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:19



SCS *Ralph J Bartlett* Wesco #166 08-15-12