

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 08/22/2012

Operator Name: JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:05
INTERNAL STANDARD	VERIFIED	19:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:05
BLANK TEST	0.00	19:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:06
BLANK TEST	0.00	19:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:07
BLANK TEST	0.00	19:08

*J. Lopez 289*

