

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/29/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:42
INTERNAL STANDARD	VERIFIED	03:42
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	03:43
BLANK TEST	0.00	03:43
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	03:44
BLANK TEST	0.00	03:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:45
BLANK TEST	0.00	03:46



SGS *Ralph Bartlett* WCSO #162 08-29-12