

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206  
 -----

Date: 09/11/2012

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:31
INTERNAL STANDARD	VERIFIED	02:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:31
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:32
BLANK TEST	0.00	02:33
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:34
BLANK TEST	0.00	02:34



SGT *Ralph J Bartlett* 09-11-12 Weso #16