

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 09/17/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:23
INTERNAL STANDARD	VERIFIED	04:24
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:24
BLANK TEST	0.00	04:25
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:25
BLANK TEST	0.00	04:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:26
BLANK TEST	0.00	04:27



565 *Ralph Bartlett* 09-17-12

WCSO #162