

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 09/24/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:34
INTERNAL STANDARD	VERIFIED	22:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:34
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:35
BLANK TEST	0.00	22:36
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	22:36
BLANK TEST	0.00	22:37



SGS Peak 1 Results WCSO #162 09-24-12