

SOLUTION CHANGE




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206  
 -----

Date: 10/01/2012

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:26
INTERNAL STANDARD	VERIFIED	02:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:27
BLANK TEST	0.00	02:28
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:28
BLANK TEST	0.00	02:29



SGT Rusky Bennett weso #162 10-01-12