

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 10/09/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12110

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 03:00 |
| INTERNAL STANDARD       | VERIFIED | 03:00 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 03:00 |
| BLANK TEST              | 0.00     | 03:01 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 03:01 |
| BLANK TEST              | 0.00     | 03:02 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 03:02 |
| BLANK TEST              | 0.00     | 03:03 |



SGT *Ralph J Bartlett* WPSO SHERIFF #162 10-09-12