

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/15/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:33
INTERNAL STANDARD	VERIFIED	02:33
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:33
BLANK TEST	0.00	02:34
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:34
BLANK TEST	0.00	02:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:35
BLANK TEST	0.00	02:36



SG Rust / Bartlett WCSO #162 10-15-12