

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/23/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:57
INTERNAL STANDARD	VERIFIED	00:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:58
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:59
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:01



SGT *Ralph J Bartlett* WCSO #162 10-23-12