

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206  
 -----

Date: 10/29/2012

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:36
INTERNAL STANDARD	VERIFIED	03:36
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:37
BLANK TEST	0.00	03:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:39



*SGT Ralph J Bartlett Weso #162 10-29-12*