

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 11/07/2012

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:23
INTERNAL STANDARD	VERIFIED	08:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	08:24
BLANK TEST	0.00	08:25
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	08:25
BLANK TEST	0.00	08:26

Handwritten signature
#248

