

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 11/13/2012

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:09
INTERNAL STANDARD	VERIFIED	14:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:10
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:11
BLANK TEST	0.00	14:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:12
BLANK TEST	0.00	14:13

Charles B. Geisler #2110

