

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 11/27/2012

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:18
INTERNAL STANDARD	VERIFIED	10:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:18
BLANK TEST	0.00	10:19
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	10:19
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:20
BLANK TEST	0.00	10:21

C. B. Geisler #241

