

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/01/2012

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:59
INTERNAL STANDARD	VERIFIED	15:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:59
BLANK TEST	0.00	16:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	16:00
BLANK TEST	0.00	16:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	16:01
BLANK TEST	0.00	16:02

