

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/13/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:37
INTERNAL STANDARD	VERIFIED	01:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:38
BLANK TEST	0.00	01:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:39
BLANK TEST	0.00	01:40
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:40
BLANK TEST	0.00	01:41



SGT Ralph J Bartlett

WCSO #162

12-13-12